

NANOCOLOR® Nitrat 250

Nitrate / Nitrato



Test 0-66
REF 985066

Methode / Method / Méthode / Método
0661 4–60 mg/L NO₃-N
0662 20–250 mg/L NO₃⁻

345 / 350 / 365 nm

Min	20–80 %		Max
4	15	49	60 mg/L NO ₃ -N
20	66	204	250 mg/L NO ₃ ⁻

