


# NANOCOLOR® Chlordioxid 5

Chlorine dioxide 5 / Dioxyde de chlore 5 / Cloro dióxido 5



Test 0-18

REF 985018

Methode / Method / Méthode / Método

0181 0.15–5.00 mg/L

530 / 540 nm

Min

0.15

20–80 %

1.12 4.03

Max

5.00 mg/L

