

REF 985049

Test 0-49 03.23

NANOCOLOR® Silver 3

en

Method:

Silver ions react with an indicator to form a blue dye.

Range:	Tube test 0.20–3.00 mg/L Ag ⁺	50 mm semi-micro cuvette 0.08–0.50 mg/L Ag ⁺ not linear
Factor:	03.90	
Wavelength (HW = 5–12 nm):	620 nm	
Reaction time:	10 min (600 s)	
Reaction temperature:	20–25 °C	

Contents of reagent set:

20 test tubes Silver 3

1 test tube with 11 mL Silver 3 R2

1 test tube with 11 mL Silver 3 R3

Hazard warning:

This test does not contain any harmful substances which must be specially labelled as hazardous.

Interferences:

Silver compounds like silver bromide, silver chloride, silver iodide, silver cyanide or silver thiocyanate are not detected by the determination. These compounds can be determined after pretreatment with NANOCOLOR® NanOx Metal (REF 918978).

The following ions will not interfere:

< 1000 mg/L Pb²⁺, F⁻, NO₃⁻, SO₄²⁻< 500 mg/L PO₄³⁻< 200 mg/L Mn²⁺, Ni²⁺< 100 mg/L Al³⁺, Cr(III)< 50 mg/L Cd²⁺< 20 mg/L Ca²⁺, Cu²⁺, Fe³⁺, Hg²⁺, Mg²⁺, Zn²⁺

< 10 mg/L Cr(VI), Mo(VI)

The method can not be applied for the analysis of sea water.

Procedure:

Requisite accessories: piston pipette with tips

Open test tube, add

500 µL (= 0.5 mL) R2 and

4.0 mL test sample (*the pH value of the sample must be between pH 3 and 9*), close and mix.

Open test tube again, add

500 µL (= 0.5 mL) R3, close and mix.

Clean outside of test tube and measure after 10 min.

Lower silver concentrations (0.08–0.50 mg/L Ag⁺) can be determined by using 50 mm semi-micro cuvettes (REF 91950):

Test sample	Blank value
Open test tube, add 500 µL (= 0.5 mL) R2 and 5.0 mL test sample (<i>the pH value of the sample must be between pH 3 and 9</i>), close and mix. Open test tube again, add 500 µL (= 0.5 mL) R3, close and mix.	Open test tube, add 500 µL (= 0.5 mL) R2 and 5.0 mL distilled water, close and mix. Open test tube again, add 500 µL (= 0.5 mL) R3, close and mix.

Pour the contents of test tubes into 50 mm semi-micro cuvettes and measure after 10 min [method 1491].

Measurement:

For NANOCOLOR® photometers see manual, test 0-49.

Measurement when samples are colored or turbid:

For all NANOCOLOR® photometers see manual, use key for correction value.

Photometers of other manufacturers:

For other photometers check whether measurement of round glass tubes is possible. Verify factor for each type of instrument by measuring standard solutions.

