


NANOCOLOR[®] gesamt-Phosphat 50

total Phosphate / Phosphate total / Fosfato total













Test 0-79
REF 985079

Methode / Method / Méthode / Método

436 nm


0791 10.0–50.0 mg/L P
0792 30–150 mg/L PO₄³⁻
0793 23–110 mg/L P₂O₅

Min	20–80 %		Max
10.0	18.0	42.0	50.0 mg/L P
30	54	126	150 mg/L PO ₄ ³⁻



NANOCOLOR® ortho-Phosphat 50

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-79
REF 985079

Methode / Method / Méthode / Método

436 nm

0794 10.0–50.0 mg/L $\text{PO}_4\text{-P}$

0795 30–150 mg/L PO_4^{3-}

0796 23–110 mg/L P_2O_5

Min

20–80 %

Max

10.0

18.0

42.0

50.0 mg/L $\text{PO}_4\text{-P}$

30

54

126

150 mg/L PO_4^{3-}

