


NANOCOLOR® Blei 5

Lead 5/Plomb 5/Plomo 5



Test 0-09
REF 985009

Methode / Method / Méthode / Método
0091 0.10–5.00 mg/L

520 nm / 540 nm

Min	20–80 %		Max
0.10	1.08	4.02	5.00 mg/L

