


MACHEREY-NAGEL

User manual



■ NucleoSpin® RNA/DNA Buffer Set

September 2023 / Rev. 12

MACHEREY-NAGEL

www.mn-net.com



RNA and DNA purification

Protocol at a glance (Rev.12)
















NucleoSpin® RNA/miRNA/RNA Blood/ RNA Plant, NucleoSpin® RNA/Protein				NucleoSpin® RNA XS	
1 Homogenize sample		Sample		Sample	
2 Lyse cells		350 µL RA1, RAP, or RP1 3.5 µL reducing agent		100 µL RA1 2 µL TCEP	
		Mix		Mix	
3 Filtrate lysate	 	11,000 x g 1 min		11,000 x g 30 s	
				5 µL Carrier RNA	
4 Adjust RNA binding conditions		350 µL 70 % ethanol Mix		100 µL 70 % ethanol Mix	
5 Bind RNA/DNA	 	Load lysate 11,000 x g 30 s		Load lysate 11,000 x g 30 s	
A Wash silica membrane	 	1 st wash 500 µL DNA Wash		400 µL DNA Wash	NucleoSpin® RNA/DNA Buffer Set
		2 nd wash 500 µL DNA Wash		400 µL DNA Wash	
		11,000 x g 1 min		11,000 x g 1 min	
B Dry membrane		RT, 3 min		RT, 3 min	
C Elute DNA	 	100 µL DNA Elute		80 µL DNA Elute	
		11,000 x g 1 min		11,000 x g 1 min	
7 Digest DNA		95 µL DNase reaction mixture RT, 15 min		25 µL DNase reaction mixture RT, 15 min	
8 Wash and dry silica membrane	  	1 st wash 200 µL RA2		100 µL RA2	
		2 nd wash 600 µL RA3		400 µL RA3	
		3 rd wash 250 µL RA3		200 µL RA3	
		1 st and 2 nd 11,000 x g 30 s		11,000 x g 30 s	
9 Elute highly pure RNA	 	11,000 x g 2 min		11,000 x g 2 min	
		60 µL RNase-free H ₂ O		10 µL RNase-free H ₂ O	
		11,000 x g 1 min		11,000 x g 30 s	

Table of contents

1 Components	4
1.1 Set contents	4
1.2 Consumables and equipment to be supplied by user	4
1.3 About this user manual	4
2 Product description	5
2.1 The basic principle	5
2.2 Kit specifications	6
3 Storage conditions and preparation of working solutions	8
4 Safety instructions	8
4.1 Disposal	8
5 Protocol – isolation of RNA and DNA from one undivided sample	9
6 Appendix	11
6.1 Troubleshooting	11
6.2 Ordering information	12
6.3 Product use restriction / warranty	13

1 Components

1.1 Set contents

NucleoSpin® RNA/DNA Buffer Set	
REF	100 preps 740944
Buffer DNA Wash (Concentrate)*	2 × 12 mL
Buffer DNA Elute	12 mL
User manual	1

1.2 Consumables and equipment to be supplied by user

The content of this set is sufficient for 100 DNA isolations in combination with RNA isolations performed with the following kits:

NucleoSpin® RNA (REF 740955), NucleoSpin® miRNA (REF 740971), NucleoSpin® RNA Blood (REF 740200), NucleoSpin® RNA Plant (REF 740949), NucleoSpin® RNA/Protein (REF 740933), NucleoSpin® RNA XS (REF 740902), NucleoSpin® 8 RNA (REF 740698), NucleoSpin® 8 RNA Core Kit (REF 740456), NucleoSpin® 96 RNA (REF 740709), NucleoSpin® 96 RNA Core Kit (REF 740466).

Additional collection tubes are required and are not supplied (see section 6.2, ordering information).

1.3 About this user manual

It is strongly recommended reading the detailed protocol sections of this user manual if the **NucleoSpin® RNA/DNA Buffer Set** is used in combination with NucleoSpin® RNA (REF 740955), NucleoSpin® miRNA (REF 740971), NucleoSpin® RNA Blood (REF 740200), NucleoSpin® RNA Plant (REF 740949), NucleoSpin® RNA/Protein (REF 740933), NucleoSpin® RNA XS (REF 740902), NucleoSpin® 8 RNA (REF 740698), NucleoSpin® 8 RNA Core Kit (REF 740456), NucleoSpin® 96 RNA (REF 740709), or NucleoSpin® 96 RNA Core Kit (REF 740466) for the first time. Experienced users, however, may refer to the Protocol at a glance instead. The Protocol at a glance is designed to be used only as a supplemental tool for quick referencing while performing the purification procedure.

All technical literature is available on the internet at www.mn-net.com.

* For preparation of working solutions and storage conditions see section 3.

2 Product description

2.1 The basic principle

The **NucleoSpin® RNA/DNA Buffer Set** is intended to be used with one of the following RNA purification kits: NucleoSpin® RNA, NucleoSpin® miRNA, NucleoSpin® RNA Blood, NucleoSpin® RNA Plant, NucleoSpin® RNA/Protein, NucleoSpin® RNA XS, NucleoSpin® 8 RNA, NucleoSpin® 8 RNA Core Kit, NucleoSpin® 96 RNA, or NucleoSpin® 96 RNA Core Kit. The combination the **NucleoSpin® RNA/DNA Buffer Set** with either of the RNA purification kits enables the isolation of RNA and DNA from one undivided sample with one single NucleoSpin® RNA Binding Column. This patented technology enables successive elution of DNA and RNA from a NucleoSpin® Column with low salt buffer and water respectively. DNA and RNA are immediately ready for downstream applications. Samples are lysed in the lysis buffer supplied in the NucleoSpin® RNA kits (Lysis Buffer RA1, RAP, or RP1). Ethanol is added to facilitate conditions for binding of nucleic acids to the NucleoSpin® RNA Binding Column. After wash steps DNA and RNA are eluted sequentially. DNA is eluted with a low salt solution (DNA Elute) which selectively elutes DNA and keeps RNA on the column. Eluted DNA is immediately ready for downstream applications without further purification. DNA eluted with DNA Elute may readily serve as template for PCR, is restrictable with restrictions enzymes and is of high molecular weight (≥ 20 kb). A_{260}/A_{280} ratios of eluted DNA are within a range from 1.7–2.0.

After DNA elution, residual on-column-DNA is digested on the NucleoSpin® Column as described in the relating NucleoSpin® RNA protocol. After additional washing steps, pure RNA is eluted with RNase-free water. DNA elution prior to RNA elution does neither compromise RNA quality nor quantity. Sequential DNA and RNA isolation from one sample with this support set and NucleoSpin® RNA kits has been successfully performed with various sample materials (e.g., HeLa cells, pig liver, kidney and spleen, parsley leaf, maize leaf, and root).

The standard protocol (section 5) allows the purification of DNA and RNA from a variety of sample types. Suitable sample types are described in the respective user manuals of the NucleoSpin® RNA kits.

2.2 Kit specifications

Typical yields of RNA and DNA




Figure 1 DNA and RNA yield from different amounts of HeLa cells

Different amounts of HeLa cells were used as sample material. DNA and RNA were isolated with the NucleoSpin® RNA/DNA Buffer Set in combination with the NucleoSpin® RNA kit.

DNA and RNA were isolated as described in Figure 1. Obtained correlation coefficients between sample amount and RNA and DNA yield are shown in Table 1.

Table 1: Correlation between sample amount and nucleic acid yield

	$3 \times 10^4 - 5 \times 10^5$ cells	$3 \times 10^4 - 1 \times 10^6$ cells
RNA	> 0.98	> 0.98
DNA	> 0.99	> 0.95
Use	For research use only	




Figure 2 DNA and RNA yield from different amounts of mouse liver tissue

Different amounts of mouse liver tissue were used as sample material. DNA and RNA were isolated with the NucleoSpin® RNA/DNA Buffer Set in combination with the NucleoSpin® RNA kit.

DNA and RNA were isolated as described in Figure 2. Obtained correlation coefficients between sample amount and RNA and DNA yield are shown in Table 2.

Table 2: Correlation between sample amount and nucleic acid yield

	0.08 – 1.25 mg mouse liver	0.08 – 2.5 mg mouse liver	0.08 – 5 mg mouse liver
RNA	> 0.98	> 0.98	> 0.98
DNA	> 0.99	> 0.95	> 0.67

DNA size and quality

- Isolated genomic DNA is commonly of high molecular weight > 20 kb.
- DNA is commonly stable, even at 37 °C for 2 h with or without addition of a typical restriction enzyme buffer.
- DNA is digestible with restriction enzymes.
- DNA is suitable for PCR.

3 Storage conditions and preparation of working solutions

Store solutions at room temperature (15–25 °C).

- The DNA Wash solution is delivered as a concentrate. To prepare the final DNA Wash solution, add four volumes of ethanol (50 %) to the DNA Wash Concentrate (add 48 mL 50 % ethanol to 12 mL DNA Wash Concentrate).
- Due to its composition DNA Elute (DNA elution buffer) does not inhibit DNases, i.e., DNA Elute does not contain substances (e.g., EDTA) to complex divalent cations. Therefore, make sure not to contaminate DNA Elute with DNases!
- Further, due to its composition, DNA Elute does not inhibit microbial growth. Therefore, make sure not to contaminate DNA Elute with any source of microbial contaminants.

NucleoSpin® RNA/DNA Buffer Set	
REF	100 preps 740944
Buffer DNA Wash (Concentrate)	2 × 12 mL Add 48 mL ethanol (50 %) to each bottle

4 Safety instructions

The **NucleoSpin® RNA/DNA Buffer Set** is intended to be used in conjunction with NucleoSpin® RNA kits. The **NucleoSpin® RNA/DNA Buffer Set** does not contain hazardous contents. However, pay attention to the safety instructions of the individual NucleoSpin® RNA kits!

4.1 Disposal

Dispose hazardous, infectious or biologically contaminated materials in a safe and acceptable manner and in accordance with all local and regulatory requirements.

5 Protocol – isolation of RNA and DNA from one undivided sample

Before starting the procedure:

- Check if Buffer DNA Wash was prepared according to section 3.
- Perform sample homogenization, cell lysis, lysate filtration, adjusting of nucleic acid binding conditions, and binding of nucleic acids to the NucleoSpin® RNA Binding Column according to the NucleoSpin® RNA, NucleoSpin® miRNA, NucleoSpin® RNA Blood, NucleoSpin® RNA Plant, NucleoSpin® RNA/Protein, NucleoSpin® RNA XS, NucleoSpin® 8 RNA, NucleoSpin® 8 RNA Core Kit, NucleoSpin® 96 RNA or NucleoSpin® 96 RNA Core Kit kit standard protocol.

Subsequent to binding of nucleic acids to the column continue as follows with step A (the membrane desalting step of the individual NucleoSpin® RNA protocols is replaced by steps A–C):

A Wash silica membrane

1st wash

Add **500 µL DNA Wash** to the NucleoSpin® RNA Binding Column and centrifuge for **1 min** at **11,000 x g**. Discard flowthrough and reuse Collection Tube.

If using NucleoSpin® RNA XS add only 400 µL DNA Wash.

The DNA Wash solution is used instead of MDB (Membrane Desalting Buffer) from the NucleoSpin® RNA kits. MDB will not be used in this procedure.



**+ 500 µL
DNA Wash**
**11,000 x g,
1 min**

2nd wash

Add again **500 µL DNA Wash** and centrifuge **1 min** at **11,000 x g**. Discard Collection Tube with flowthrough.

If using NucleoSpin® RNA XS add only 400 µL DNA Wash.



**+ 500 µL
DNA Wash**
**11,000 x g,
1 min**

B Dry membrane

Insert the NucleoSpin® RNA Binding Column into a new 1.5 mL microcentrifuge tube (not supplied). Open the lid of the NucleoSpin® RNA Binding Column and let it stand for 3 minutes.

The procedure ensures complete removal of ethanol from the column.

**Incubate
for 3 min**

C Elute DNA

Add **100 µL DNA Elute** (DNA elution buffer) directly onto the membrane and incubate 1 min. Elute the DNA by centrifuging for **1 min** at **11,000 x g**.



**Add 100 µL
DNA Elute**

If using NucleoSpin® RNA XS add only 80 µL DNA Elute for elution.

The temperature of the DNA Elute solution shall not exceed 30 °C, otherwise RNA will partly elute with the DNA Elute solution. DNA Elute solution may stay for 1 min up to 15 min on the column before DNA is eluted. A 1–5 min incubation time is recommended. Eluted DNA is immediately ready for downstream applications without further purification.



**11,000 x g,
1 min**

Proceed with the digestion of residual on-column DNA according to the individual NucleoSpin® RNA protocols (step: Digest DNA): Add DNase reaction mixture onto the column and perform all subsequent steps as described in the **NucleoSpin® RNA**, **NucleoSpin® miRNA**, **NucleoSpin® RNA Blood**, **NucleoSpin® RNA Plant**, **NucleoSpin® RNA/Protein**, **NucleoSpin® RNA XS**, **NucleoSpin® 8 RNA**, **NucleoSpin® 8 RNA Core Kit**, **NucleoSpin® 96 RNA**, or **NucleoSpin® 96 RNA Core Kit** protocol.

6 Appendix

6.1 Troubleshooting

Problem	Possible cause and suggestions
DNA is contaminated with RNA	<p><i>Buffer temperature</i></p> <ul style="list-style-type: none"> DNA elution buffer DNA Elute exceeded 30 °C during application. Use DNA Elute with a temperature preferentially of 15–25 °C.
DNA yield lower than RNA yield	<p><i>Sample material</i></p> <ul style="list-style-type: none"> DNA and RNA yield depend very much on sample material. Ratio of RNA yield to DNA yield may vary from approximately 1–20.
DNA degrades upon storage	<p><i>DNase contamination</i></p> <ul style="list-style-type: none"> DNA elution buffer DNA Elute does not contain divalent cations complexing substances (e.g., EDTA). Therefore, DNA is not protected against DNases. Keep DNA Elute solution clean and avoid any contamination. As a precaution, keep DNA on ice for short term or at -20 °C for long term storage Some sample materials may contain remaining DNase traces that are not sufficiently washed away by the standard procedure. Perform a wash step of the column with Buffer RA2 after loading the lysate onto the column and before starting the washing steps with DNA Wash solution: Add 500 µL Buffer RA2 onto the column, centrifuge 1 min at 11,000 x g and continue with DNA Wash washing steps.
Low RNA yield or quality	<p><i>See general protocol</i></p> <ul style="list-style-type: none"> See troubleshooting section of individual NucleoSpin® protocols. Check if Wash Buffer RA3 has been equilibrated to room temperature before use. Washing at lower temperatures lowers efficiency of salt removal by Wash Buffer RA3.
Suboptimal performance of DNA in downstream applications	<p><i>Divalent cations</i></p> <ul style="list-style-type: none"> Eluted DNA contains small amounts of divalent cations. If the downstream application comprises for example 50 % DNA eluate of the final reaction volume the divalent cations introduced into the reaction by the DNA eluate may alter the performance. Decrease the divalent cation concentration of the reaction by 1–5 mM for compensation.
Low DNA yield for large sample amounts	<p><i>Sample amount too large</i></p> <ul style="list-style-type: none"> Depending on the type of sample and its DNA content, DNA yield may not increase proportional with increased sample amount. Sample amounts larger than for example 5 mg tissue or 10⁶ cultured cells may yield less DNA than smaller sample amounts. Use smaller sample to ensure good correlation between sample amount and DNA yield.

6.2 Ordering information

Product	REF	Pack of
NucleoSpin® RNA/DNA Buffer Set	740944	100 preps
NucleoSpin® RNA	740955.10 / .50 / .250	20 / 50 / 250 preps
NucleoSpin® miRNA	740971.10 / .50 / .250	10 / 50 / 250 preps
NucleoSpin® RNA Blood	740200.10 / .50	10 / 50 preps
NucleoSpin® RNA Plant	740949.10 / .50 / .250	10 / 50 / 250 preps
NucleoSpin® RNA/Protein	740933.10 / .50 / .250	10 / 50 / 250 preps
NucleoSpin® RNA XS	740902.10 / .50 / .250	10 / 50 / 250 preps
NucleoSpin® TriPrep*	740666.10 / .50 / .250	10 / 50 / 250 preps
NucleoSpin® 8 RNA	740698 / .5	12 × 8 / 60 × 8 preps
NucleoSpin® 8 RNA Core Kit	40456.4	48 × 8 preps
NucleoSpin® 96 RNA	740709.2 / .4 / .24	2 × 96 / 4 × 96 / 24 × 96 preps
NucleoSpin® 8 RNA Core Kit	740466.4	4 × 96 preps
Buffer RA1	740961	50 mL
Buffer RA1	740961.500	500 mL
Buffer RP1	740934.50	50 mL
Buffer RP1	740934.500	500 mL
rDNase Set	740963	1 set
NucleoSpin® Filters	740606	50
NucleoSpin® 96 RNA Filter Plate	740711	4 plates
Collection Tubes (2 mL)	740600	1000

Visit **www.mn-net.com** for more detailed product information.

6.3 Product use restriction / warranty

All MACHEREY-NAGEL products are designed for their intended use only. They are not intended to be used for any other purpose. The description of the intended use of the products can be found in the original MACHEREY-NAGEL product leaflets. Before using our products, please observe the instructions for use and the safety instructions from the respective Material Safety Data Sheet of the product.

This MACHEREY-NAGEL product is carrying documentation stating specifications and other technical information. MACHEREY-NAGEL warrants to meet the stated specifications. The provided warranty is limited to the data specifications and descriptions as given in the original MACHEREY-NAGEL literature. No other statements or representations, written or oral, by MACHEREY-NAGEL's employees, agents or representatives, except written statements signed by a duly authorized officer of MACHEREY-NAGEL are authorized. They should not be relied upon by the customer and are not a part of a contract of sale or of this warranty.

Liability for all possible damages that occur in any connection with our products is limited to the utmost minimum as stated in the general business terms and conditions of MACHEREY-NAGEL in their latest edition which can be taken from the company's website. MACHEREY-NAGEL does not assume any further warranty.

Products and their application are subject to change. Therefore, please contact our Technical Service Team for the latest information on MACHEREY-NAGEL products. You may also contact your local distributor for general scientific information. Descriptions in MACHEREY-NAGEL literature are provided for informational purposes only.


Last updated: 08/2022, Rev. 04

Please contact:

MACHEREY-NAGEL GmbH & Co. KG

Tel.: +49 24 21 969-333

support@mn-net.com



Plasmid DNA

Clean up

RNA

DNA


Viral RNA and DNA

Protein

High throughput

Accessories

Auxiliary tools



www.mn-net.com

MACHEREY-NAGEL



A031003/093x.x



MACHEREY-NAGEL GmbH & Co. KG · Valencienner Str. 11 · 52355 Düren · Germany

DE +49 24 21 969-0 info@mn-net.com

CH +41 62 388 55 00 sales-ch@mn-net.com

FR +33 388 68 22 68 sales-fr@mn-net.com

US +1 888 321 62 24 sales-us@mn-net.com